

**INSTRUCTIONS FOR USE****SKIRROW ANTIMICROBIC SUPPLEMENT****Freeze-dried selective supplement****1 - INTENDED USE**

In vitro diagnostic. Mixture of antimicrobials to be added to a medium base for the isolation of *Campylobacter jejuni* and *Campylobacter coli* from clinical and other specimens.

2 - COMPOSITION - VIAL CONTENTS FOR 500 ML OF MEDIUM

Polymyxin B	1250 IU
Trimethoprim	2.5 mg
Vancomycin	5.0 mg

3 - PRINCIPLE OF THE METHOD AND EXPLANATION OF THE PROCEDURE

Skirrow Antimicrobial Supplement is a freeze-dried mixture of antimicrobials to be used as a supplement of Columbia Agar Base. The complete medium Skirrow Medium, originally developed in 1977 by Skirrow,¹ is a selective medium for the isolation of *Campylobacter jejuni* and *Campylobacter coli* from clinical and non-clinical specimens. Vancomycin inhibits Gram-positives bacteria, and trimethoprim and polymyxin B inhibit many Gram-negative organisms

4- DIRECTIONS

Aseptically reconstitute the content of one vial with 5 mL of sterile purified water and mix gently to dissolve.

Prepare 500 mL of Columbia Agar Base (401136) autoclaved at 121°C for 15 minutes and cooled to 47-50°C. Add the supplement together with 25 mL of lysed horse blood. Mix well and pour into sterile Petri dishes.

5 - PHYSICAL CHARACTERISTICS

Freeze-dried supplement appearance	short, dense, yellow pastille
Reconstituted supplement appearance	limpid, colourless

6 - MATERIALS PROVIDED - PACKAGING

Product	Type	REF	Pack
Skirrow Antimicrobial Supplement	Freeze-dried supplement	4240016	10 vials, each for 500 mL of medium

7 - MATERIALS REQUIRED BUT NOT PROVIDED

Columbia Agar Base (401136), autoclave, water-bath, incubator and laboratory equipment as required, sterile loops and swabs, Petri dishes, Erlenmeyer flasks, ancillary culture media and reagents for the identification of the colonies. Materials for incubation in a microaerophilic atmosphere.

8 - SPECIMENS

Skirrow medium is intended for the bacteriological processing of clinical specimens such as faeces and rectal swab^{2,3} and non-clinical specimens such as food and animal feeding stuffs. Good laboratory practices for collection, transport and storage of clinical specimens should be applied. Collect specimens before antimicrobial therapy where possible.

9 - TEST PROCEDURE

Allow plates to come to room temperature and to dry the surface of the medium.

Allow plates to come to room temperature and to dry the surface of the medium.

- Solid faeces: faeces may be diluted 1:4 in sterile saline solution or 0.1% peptone water. It has been shown that dilution significantly reduces the amount of competing flora without compromising isolation of low numbers of pathogens.³ Inoculate 3-5 drops on the medium surface.
- Liquid stool: inoculate 3 drops on the medium surface.
- Rectal swabs: roll the swab over a small area of the surface at the edge; then streak from this inoculated area.

For all type of specimens, streak with a loop over the four quadrants of the plate to obtain well isolated colonies, ensuring that sections 1 and 4 do not overlap.

Incubate in a microaerobic atmosphere consisting approximately of 5% O₂, 10% CO₂, and 85% N₂, at 39-42°C for 40-48 hours.²

10 - READING AND INTERPRETATION

After incubation, observe the bacterial growth and record the specific morphological and chromatic characteristics of the colonies.

Colonies of *C. jejuni* appear non-haemolytic, usually are grey/white or creamy grey in colour, swarming and moist in appearance. They may appear as a layer of growth over the surface of the agar. Colonies are usually non-pigmented.

Campylobacter species are oxidase positive. If a colony phenotypically resembling *Campylobacter* species is oxidase negative, subculture to blood agar and retest after 24hr incubation.⁴

The presumptive identification of thermophilic and enteropathogenic *Campylobacter* can be done on the basis of oxidase test (+) and the characteristic motility. For a complete explanation of the identification criteria and methods, refer to the quoted reference.⁴

Plates examined after 24 h of incubation should be rapidly examined and re-incubated under microaerophilic conditions to maintain viability of the strains most sensitive to oxygen.

11 - USER QUALITY CONTROL

All manufactured lots of the product are released for sale after the Quality Control has been performed to check the compliance with the specifications. However, the end user can perform its own Quality Control testing in accordance with the local applicable regulations, in





compliance with accreditation requirements and the experience of the Laboratory. Here below are listed some test strains useful for the quality control.

CONTROL STRAINS	INCUBATION T° / T / ATM	EXPECTED RESULTS
<i>C.jejuni</i> ATCC 33291	39-42°C / 40-48h / M	good growth
<i>C.coli</i> ATCC 43478	39-42°C / 40-48h / M	good growth
<i>E.coli</i> ATCC 25922	39-42°C / 40-48h / M	partially or totally inhibited
<i>S.aureus</i> ATCC 25923	39-42°C / 40-48h / M	inhibited

M: microaerobic incubation; ATCC is a trademark of American Type Culture Collection

12 - PERFORMANCES CHARACTERISTICS

Prior to release for sale a representative sample of all lots of Skirrow Antimicrobial Supplement added to dehydrated Blood Agar Base n° 2 Base together with 5% of lysed horse blood, is tested for productivity and selectivity by comparing the results with previously approved Reference Batch.

Productivity is tested by a semi-quantitative test with the target strains *C.jejuni* ATCC 33291 and *C.coli* ATCC 43478 ; plates are inoculated with decimal dilutions in saline of the colonies suspensions and incubated in a microaerophilic atmosphere at 39-42°C for 24 hours. Both strains show a good growth.

Selectivity is evaluated with modified Miles-Misra surface drop method by inoculating the plates with suitable decimal dilutions in saline of a 0.5 McFarland suspension of the non-target strains *E.coli* ATCC 25922, *P.aeruginosa* ATCC 27853 *S.aureus* ATCC 25923, *E.faecalis* ATCC 19433, *P.rettgeri* ATCC 39944, *P.mirabilis* ATCC 12453, *C.albicans* ATCC 60193. The growth of *E.faecalis*, *S.aureus* and *P.mirabilis* is totally inhibited while the growth of other non-target strains is partially inhibited after incubation at 39-42°C for 48 hours in a microaerophilic atmosphere.

13 - LIMITATIONS OF THE METHOD

- The most numerous contaminants found in the Skirrow medium are *Enterobacteriaceae*, which are resistant to antimicrobials.
- To achieve the highest yield of *Campylobacter* from stool samples, a combination of media that includes Skirrow medium and a second selective medium, based on a different selective system, appears to be the optimal method (e.g., Karmali medium).⁴
- The clinical advantage of enrichment broths formulated to enhance the recovery of *Campylobacter* has not been studied adequately.² Enrichment seems not to be necessary for samples collected in the acute campylobacteriosis phase, while *Campylobacter* recovery increases in asymptomatic patients, in studies involving low numbers of the target organism, in samples not readily sent to the laboratory and in samples taken in the convalescence phase after an episode of diarrhea.^{5,6}
- Even if the microbial colonies on the plates are differentiated on the basis of their morphological and chromatic characteristics, it is recommended that biochemical, immunological, molecular, or mass spectrometry testing be performed on isolates, from pure culture, for complete identification. If relevant, perform antimicrobial susceptibility testing.
- The culture medium and the supplement are intended as an aid in the diagnosis of infectious diseases; the interpretation of the results must be made considering the patient's clinical history, the origin of the sample and the results of other diagnostic tests.

14 - PRECAUTIONS AND WARNINGS

- The supplement is a qualitative *in vitro* diagnostic, for professional use only; it must be used by adequately trained and qualified laboratory personnel, observing approved biohazard precautions and aseptic techniques.
- Antibiotics containing supplements must be handled with suitable protection; consult the Safety Data Sheet before use.
- The supplement and the medium base shall be used in association according to the directions described above.
- Apply Good Manufacturing Practice in the preparation process of plated media.
- The supplement is sterilized by membrane filtration.
- Be careful when opening the metal ring to avoid injury.
- All laboratory specimens should be considered infectious.
- The laboratory area must be controlled to avoid contaminants such as medium powder and supplements or microbial agents.
- Sterilize all biohazard waste before disposal. Dispose the unused supplements and the sterilized media inoculated with samples or microbial strains in accordance with current local legislation.
- Do not use the supplement as active ingredients for pharmaceutical preparations or as production materials intended for human and animal consumption.
- The Certificates of Analysis and the Safety Data Sheet of the product are available on the website www.biolifeitaliana.it.
- Notify Biolife Italiana Srl (complaint@biolifeitaliana.it) and the relevant Authorities of any serious incident occurring in connection with the use of the *in vitro* diagnostic.
- The information provided in this document has been defined to the best of our knowledge and ability and represents a guideline for the proper use of the product but without obligation or liability. In all cases existing local laws, regulations and standard procedures must be observed for the examination of samples collected from human and animal organic districts, for environmental samples and for products intended for human or animal consumption. Our information does not relieve our customers from their responsibility for checking the suitability of our product for the intended purpose.

15 - STORAGE CONDITIONS AND SHELF LIFE

Upon receipt, store the product in the original package at 2-8°C away from direct light. If properly stored, the product may be used up to the expiry date printed on the label; do not use beyond this date. Once the vial has been opened and the lyophilized product has been reconstituted, the resulting solution should be used immediately. Before use, examine the lyophilized and reconstituted product and discard if there are obvious signs of deterioration (e.g., contamination, atypical colour or other abnormal characteristics). The user is responsible for the manufacturing and quality control processes of plated media and the validation of their shelf life, according to the applied storage conditions (temperature and packaging).

16 - REFERENCES

1. Skirrow, M. B. 1977. *Campylobacter* enteritis: a "new" disease. *Br. Med. J.* 2:9-11.
2. Fitzgerald C, Nachamkin I. *Campylobacter* and *Arcobacter*. In Jorgensen JH, Carrol KC, Funke G et al. editors. *Manual of clinical microbiology*, 11th ed. Washington, DC: American Society for Microbiology; 2015. p.998
3. Public Health England. Investigation of Faecal Specimens for Enteric Pathogens. ID30. Issue 8.1. 2014
4. Public Health England. Identification of *Campylobacter* species. ID23. Issue 3.1. 2018





5. Endtz HP, Ruijs GJ, et al. Comparison of six media including a semisolid agar for the isolation of various *Campylobacter* species from stool specimens. *J Clin Microbiol* 1991; 29:1007
6. Bolton FJ, Robertson L. A selective medium for isolating *Campylobacter jejuni/coli*. *J Clin Pathol* 1982; 35:462
7. Hutchinson DN, Bolton FJ. Is enrichment culture necessary for the isolation of *Campylobacter jejuni* from faeces? *J Clin Pathol* 1983; 36:1350-1352

SKIRROW ANTIMICROBIC SUPPLEMENT 4240016

SDS rev 5

Regulation (EU) 2020/878

Contains: Vancomycin HCl**Classification according to Regulation (EC) No 1272/2008**

Skin sensitization, category 1

H317

May cause an allergic skin reaction

Labelling according Regulation (EC) No 1272/2008

Pictogram



Signal word

Warning

Hazard statements:

H317 May cause an allergic skin reaction.

Precautionary statements:

P280 Wear protective gloves.

P261 Avoid breathing dust / fume / gas / mist / vapours / spray.

P333+P313 If skin irritation or rash occurs: Get medical advice / attention.

P362+P364 Take off contaminated clothing and wash it before reuse.

TABLE OF APPLICABLE SYMBOLS

REF or REF Catalogue number	LOT Batch code	IVD <i>In vitro</i> Diagnostic Medical Device	Manufacturer	This side up	
Temperature limitation	Content sufficient for <n> tests	Consult Instructions for Use	Use by	Keep away from direct light	Fragile

REVISION HISTORY

Version	Description of changes	Date
Revision 1	Updated layout and content	2022/02
Revision 2	Removal of obsolete classification	2023/04

Note: minor typographical, grammatical, and formatting changes are not included in the revision history.

